

## EC Declaration of Conformity

In accordance with the Machinery Directive 2006/42/EC Annex II Letter 1.A

**Manufacturer:** Köttermann GmbH  
Industriestraße 2  
31311 Uetze  
Germany

**Product designation:** Pharmacy hood, Line EXPLORIS

**Types:** 2-441-CKO, 2-441-CLO

**Serial Number:** 2-441-019610 to 2-441-021600

The manufacturer hereby declares that the product fulfils all the relevant provisions of the directive mentioned above. Furthermore the manufacturer declares the conformity with other applied directives and relevant provisions with which the product complies.

**The following EC Directives were also applied:**

2014/30/EU Electromagnetic Compatibility  
2014/35/EU Low Voltage

**The following harmonised standards were applied:**

EN ISO 12100:2010 Safety of machinery - General principles for design - Risk assessment and risk reduction

DIN EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements


DIN EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

**The following national and international standards were applied:**

DIN 12924-4:2012 Laboratory furniture – Fume cupboards – Part 4: Fume cupboards for pharmacies

**Marking:** 

Person that is authorised to compile the relevant technical documentation:  
Köttermann GmbH, Mike Dithmers, Industriestraße 2, 31311 Uetze, Germany



Kazim Doyuran, Managing Director  
Uetze/Hänigsen, 11.04.2019